

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061319\
 Data File : VN056285.D
 Acq On : 13 Jun 2019 16:42
 Operator : JC/SP
 Sample : K3350-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OUTFALL001-190611

Quant Time: Jun 14 02:00:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	377121	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	628444	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	611746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	177934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	208022	55.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.10%	
35) Dibromofluoromethane	7.59	113	193938	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
50) Toluene-d8	10.09	98	770497	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
62) 4-Bromofluorobenzene	12.40	95	241442	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.83	96	37157	9.319	ug/l	90
44) Trichloroethene	8.83	130	1513	0.321	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061319\
Data File : VN056285.D
Acq On : 13 Jun 2019 16:42
Operator : JC/SP
Sample : K3350-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OUTFALL001-190611

Quant Time: Jun 14 02:00:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

