

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061319\
 Data File : VN056288.D
 Acq On : 13 Jun 2019 17:54
 Operator : JC/SP
 Sample : K3345-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-MW-5 061119

Quant Time: Jun 14 02:07:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	339592	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	561265	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	544888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	156049	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	191394	56.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.52%	
35) Dibromofluoromethane	7.59	113	173501	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
50) Toluene-d8	10.09	98	688031	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
62) 4-Bromofluorobenzene	12.40	95	213715	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
Target Compounds						
16) Acetone	3.82	43	6097	3.781	ug/l	90
30) Chloroform	7.37	83	4048	0.750	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061319\
Data File : VN056288.D
Acq On : 13 Jun 2019 17:54
Operator : JC/SP
Sample : K3345-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-MW-5_061119

Quant Time: Jun 14 02:07:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

