

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072802.D
 Acq On : 13 Jun 2022 13:04
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 14 05:20:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	256728	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	465149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	402409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	143594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	207859	28.917	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	57.840%#
35) Dibromofluoromethane	8.022	113	153870	33.215	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	66.420%#
50) Toluene-d8	10.439	98	512038	27.522	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	55.040%#
62) 4-Bromofluorobenzene	12.727	95	195221	29.693	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	59.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072802.D
Acq On : 13 Jun 2022 13:04
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 14 05:20:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

