

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072812.D
 Acq On : 13 Jun 2022 22:33
 Operator : JC\MD
 Sample : PB145459ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145459ZHE#03

Quant Time: Jun 14 05:23:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	157664	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	323193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	301228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	101485	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	247030	55.961	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.920%
35) Dibromofluoromethane	8.022	113	167091	51.911	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.820%
50) Toluene-d8	10.439	98	589273	45.586	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	91.180%
62) 4-Bromofluorobenzene	12.727	95	205294	44.941	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.880%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	51577	25.729	ug/l	# 86
18) Methyl Acetate	4.857	43	7201	1.367	ug/l	# 80
20) Methylene Chloride	5.098	84	8921	1.901	ug/l	# 91
25) 2-Butanone	7.339	43	20764	6.604	ug/l	# 83
43) Isopropyl Acetate	8.563	43	74573	7.030	ug/l	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072812.D
Acq On : 13 Jun 2022 22:33
Operator : JC\MD
Sample : PB145459ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145459ZHE#03

Quant Time: Jun 14 05:23:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

