

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072813.D
 Acq On : 13 Jun 2022 22:57
 Operator : JC\MD
 Sample : PB145459ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145459ZHE#04

Quant Time: Jun 14 05:23:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	162308	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	336539	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	309907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	105513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	242118	53.278	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.560%
35) Dibromofluoromethane	8.022	113	167175	49.878	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.760%
50) Toluene-d8	10.439	98	589773	43.815	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.640%
62) 4-Bromofluorobenzene	12.727	95	205909	43.288	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.580%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	51514	24.962	ug/l	# 88
18) Methyl Acetate	4.863	43	8597	1.585	ug/l	# 62
20) Methylene Chloride	5.092	84	8925	1.806	ug/l	# 84
25) 2-Butanone	7.339	43	19537	6.036	ug/l	99
43) Isopropyl Acetate	8.563	43	75652	6.849	ug/l	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072813.D
Acq On : 13 Jun 2022 22:57
Operator : JC\MD
Sample : PB145459ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145459ZHE#04

Quant Time: Jun 14 05:23:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

