

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072815.D
 Acq On : 13 Jun 2022 23:46
 Operator : JC\MD
 Sample : PB145459ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145459ZHE#06

Quant Time: Jun 14 05:23:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	161277	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.969	114	337832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	309312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	103540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	246477	54.584	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.160%	
35) Dibromofluoromethane	8.022	113	169710	50.440	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.880%	
50) Toluene-d8	10.439	98	579052	42.854	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 85.700%	
62) 4-Bromofluorobenzene	12.727	95	201817	42.265	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 84.540%	
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	50126	24.445	ug/l	99
18) Methyl Acetate	4.869	43	8032	1.490	ug/l #	75
20) Methylene Chloride	5.092	84	9260	1.951	ug/l #	83
25) 2-Butanone	7.339	43	18420	5.727	ug/l	98
43) Isopropyl Acetate	8.563	43	103576	9.341	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072815.D
Acq On : 13 Jun 2022 23:46
Operator : JC\MD
Sample : PB145459ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145459ZHE#06

Quant Time: Jun 14 05:23:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

