

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072817.D
 Acq On : 14 Jun 2022 00:35
 Operator : JC\MD
 Sample : PB145459ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145459ZHE#08

Quant Time: Jun 14 05:23:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	157327	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	326454	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	304882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	105002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	237228	53.855	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.720%	
35) Dibromofluoromethane	8.016	113	161060	49.538	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.080%	
50) Toluene-d8	10.439	98	551257	42.219	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 84.440%	
62) 4-Bromofluorobenzene	12.727	95	194836	42.225	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 84.460%	
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	33504	16.749	ug/l	# 87
18) Methyl Acetate	4.857	43	8744	1.663	ug/l	# 86
20) Methylene Chloride	5.093	84	23105	7.309	ug/l	91
25) 2-Butanone	7.334	43	8094	2.580	ug/l	# 75
43) Isopropyl Acetate	8.563	43	68809	6.422	ug/l	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072817.D
Acq On : 14 Jun 2022 00:35
Operator : JC\MD
Sample : PB145459ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145459ZHE#08

Quant Time: Jun 14 05:23:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

