

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072820.D
 Acq On : 14 Jun 2022 01:48
 Operator : JC\MD
 Sample : N3304-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-10

Quant Time: Jun 14 05:24:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	142280	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	295019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	269455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	94199	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	231101	58.013	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 116.020%	
35) Dibromofluoromethane	8.022	113	156459	53.250	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.500%	
50) Toluene-d8	10.439	98	520827	44.139	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 88.280%	
62) 4-Bromofluorobenzene	12.727	95	186067	44.622	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 89.240%	
Target Compounds						
						Qvalue
16) Acetone	4.287	43	69463	38.397	ug/l	91
18) Methyl Acetate	4.851	43	24340	5.119	ug/l	91
20) Methylene Chloride	5.093	84	42237	16.295	ug/l #	85
25) 2-Butanone	7.334	43	20036	7.061	ug/l	90
39) Methylcyclohexane	9.457	83	16519	3.560	ug/l	95
43) Isopropyl Acetate	8.557	43	44899	4.637	ug/l #	76
73) Isopropylbenzene	12.575	105	14925	1.554	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072820.D
Acq On : 14 Jun 2022 01:48
Operator : JC\MD
Sample : N3304-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-10

Quant Time: Jun 14 05:24:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

