

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072821.D
 Acq On : 14 Jun 2022 02:12
 Operator : JC\MD
 Sample : N3304-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-10-G

Quant Time: Jun 14 05:24:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	148391	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	307996	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	292495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	106070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	217691	52.396	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.800%
35) Dibromofluoromethane	8.016	113	150284	48.993	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.980%
50) Toluene-d8	10.439	98	527379	42.811	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	85.620%
62) 4-Bromofluorobenzene	12.727	95	194670	44.718	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.440%
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	79961	42.380	ug/l	# 87
18) Methyl Acetate	4.851	43	7940	1.601	ug/l	# 82
20) Methylene Chloride	5.093	84	39140	14.312	ug/l	# 80
25) 2-Butanone	7.334	43	27595	9.325	ug/l	# 81
43) Isopropyl Acetate	8.557	43	56111	5.550	ug/l	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072821.D
Acq On : 14 Jun 2022 02:12
Operator : JC\MD
Sample : N3304-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-10-G

Quant Time: Jun 14 05:24:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

