

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072824.D
 Acq On : 14 Jun 2022 03:25
 Operator : JC\MD
 Sample : N3304-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-6-G

Quant Time: Jun 14 05:25:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	143626	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	299651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	271680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	96980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	230084	57.216	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 114.440%	
35) Dibromofluoromethane	8.016	113	156308	52.376	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.760%	
50) Toluene-d8	10.439	98	517411	43.171	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 86.340%	
62) 4-Bromofluorobenzene	12.727	95	186650	44.070	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.140%	
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	69529	38.074	ug/l	92
18) Methyl Acetate	4.857	43	11197	2.333	ug/l	91
20) Methylene Chloride	5.087	84	29015	10.613	ug/l #	82
43) Isopropyl Acetate	8.563	43	50426	5.127	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072824.D
Acq On : 14 Jun 2022 03:25
Operator : JC\MD
Sample : N3304-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-6-G

Quant Time: Jun 14 05:25:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

