

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072805.D
 Acq On : 13 Jun 2022 19:18
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 14 05:21:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.781	0.694	11.1	94	0.00
3 P	Chloromethane	0.901	0.753	16.4	99	0.00
4 C	Vinyl Chloride	0.899	0.833	7.3#	102	0.00
5 T	Bromomethane	0.361	0.366	-1.4	109	0.00
6 T	Chloroethane	0.651	0.607	6.8	105	0.00
7 T	Trichlorofluoromethane	1.396	1.288	7.7	97	0.00
8 T	Diethyl Ether	0.547	0.533	2.6	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.771	0.699	9.3	96	0.00
10 T	Methyl Iodide	0.531	0.572	-7.7	105	0.00
11 T	Tert butyl alcohol	1.376	1.422	-3.3	105	0.00
12 CM	1,1-Dichloroethene	0.695	0.623	10.4#	98	0.00
13 T	Acrolein	0.072	0.059	18.1	91	0.00
14 T	Allyl chloride	1.616	1.498	7.3	100	0.00
15 T	Acrylonitrile	0.651	0.690	-6.0	107	0.00
16 T	Acetone	0.636	0.659	-3.6	110	0.00
17 T	Carbon Disulfide	1.665	1.322	20.6	97	0.00
18 T	Methyl Acetate	1.671	1.679	-0.5	105	0.00
19 T	Methyl tert-butyl Ether	3.240	3.143	3.0	101	0.00
20 T	Methylene Chloride	1.008	0.895	11.2	104	0.00
21 T	trans-1,2-Dichloroethene	0.779	0.700	10.1	102	0.00
22 T	Diisopropyl ether	3.384	3.449	-1.9	105	0.00
23 T	Vinyl Acetate	3.061	3.129	-2.2	104	0.00
24 P	1,1-Dichloroethane	1.770	1.714	3.2	103	0.00
25 T	2-Butanone	0.997	1.051	-5.4	107	0.00
26 T	2,2-Dichloropropane	1.638	1.329	18.9	89	0.00
27 T	cis-1,2-Dichloroethene	0.987	0.950	3.7	101	0.00
28 T	Bromochloromethane	0.852	0.899	-5.5	106	0.00
29 T	Tetrahydrofuran	0.618	0.657	-6.3	107	0.00
30 C	Chloroform	1.838	1.824	0.8#	103	0.00
31 T	Cyclohexane	1.481	1.351	8.8	100	0.00
32 T	1,1,1-Trichloroethane	1.625	1.549	4.7	100	0.00
33 S	1,2-Dichloroethane-d4	1.400	1.415	-1.1	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.498	0.516	-3.6	107	0.00
36 T	1,1-Dichloropropene	0.694	0.647	6.8	100	0.00
37 T	Ethyl Acetate	1.060	1.101	-3.9	108	0.00
38 T	Carbon Tetrachloride	0.727	0.710	2.3	103	0.00
39 T	Methylcyclohexane	0.786	0.713	9.3	97	0.00
40 TM	Benzene	2.056	2.009	2.3	102	0.00
41 T	Methacrylonitrile	0.508	0.571	-12.4	105	0.00
42 TM	1,2-Dichloroethane	0.888	0.881	0.8	103	0.00
43 T	Isopropyl Acetate	1.641	1.684	-2.6	104	0.00
44 TM	Trichloroethene	0.468	0.449	4.1	99	0.00
45 C	1,2-Dichloropropane	0.578	0.570	1.4#	104	0.00
46 T	Dibromomethane	0.388	0.392	-1.0	103	0.00
47 T	Bromodichloromethane	0.806	0.824	-2.2	103	0.00
48 T	Methyl methacrylate	0.762	0.826	-8.4	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
 Data File : VN072805.D
 Acq On : 13 Jun 2022 19:18
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 14 05:21:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 13 17:28:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.013	0.014	-7.7	105	0.00
50 S	Toluene-d8	2.000	2.084	-4.2	106	0.00
51 T	4-Methyl-2-Pentanone	1.089	1.183	-8.6	106	0.00
52 CM	Toluene	1.305	1.284	1.6#	101	0.00
53 T	t-1,3-Dichloropropene	0.856	0.867	-1.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.908	0.888	2.2	99	0.00
55 T	1,1,2-Trichloroethane	0.553	0.548	0.9	101	0.00
56 T	Ethyl methacrylate	0.966	1.008	-4.3	102	0.00
57 T	1,3-Dichloropropane	0.993	1.009	-1.6	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.408	0.449	-10.0	101	0.00
59 T	2-Hexanone	0.806	0.895	-11.0	105	0.00
60 T	Dibromochloromethane	0.575	0.588	-2.3	100	0.00
61 T	1,2-Dibromoethane	0.558	0.568	-1.8	102	0.00
62 S	4-Bromofluorobenzene	0.707	0.718	-1.6	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.383	0.378	1.3	102	0.00
65 PM	Chlorobenzene	1.318	1.291	2.0	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.503	0.519	-3.2	99	0.00
67 C	Ethyl Benzene	2.462	2.476	-0.6#	100	0.00
68 T	m/p-Xylenes	0.881	0.871	1.1	99	0.00
69 T	o-Xylene	0.886	0.895	-1.0	99	0.00
70 T	Styrene	1.411	1.460	-3.5	100	0.00
71 P	Bromoform	0.366	0.381	-4.1	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	5.096	5.286	-3.7	99	0.00
74 T	N-amyl acetate	2.444	2.816	-15.2	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.798	1.945	-8.2	100	0.00
76 T	1,2,3-Trichloropropane	1.716	1.648	4.0	100	0.00
77 T	Bromobenzene	1.022	1.030	-0.8	98	0.00
78 T	n-propylbenzene	5.675	6.003	-5.8	99	0.00
79 T	2-Chlorotoluene	3.555	3.764	-5.9	101	0.00
80 T	1,3,5-Trimethylbenzene	4.120	4.274	-3.7	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.563	0.590	-4.8	92	0.00
82 T	4-Chlorotoluene	3.312	3.375	-1.9	97	0.00
83 T	tert-Butylbenzene	3.671	3.695	-0.7	97	0.00
84 T	1,2,4-Trimethylbenzene	3.957	4.207	-6.3	97	0.00
85 T	sec-Butylbenzene	4.915	5.223	-6.3	97	0.00
86 T	p-Isopropyltoluene	3.829	4.114	-7.4	97	0.00
87 T	1,3-Dichlorobenzene	1.753	1.766	-0.7	95	0.00
88 T	1,4-Dichlorobenzene	1.730	1.714	0.9	93	0.00
89 T	n-Butylbenzene	3.247	3.431	-5.7	95	0.00
90 T	Hexachloroethane	0.819	0.853	-4.2	98	0.00
91 T	1,2-Dichlorobenzene	1.736	1.767	-1.8	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.394	0.440	-11.7	99	0.00
93 T	1,2,4-Trichlorobenzene	0.788	0.843	-7.0	94	0.00
94 T	Hexachlorobutadiene	0.379	0.382	-0.8	95	0.00
95 T	Naphthalene	3.108	3.512	-13.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072805.D
Acq On : 13 Jun 2022 19:18
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 14 05:21:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 13 17:28:28 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.820	0.897	-9.4	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6