



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072795.D
Acq On : 13 Jun 2022 09:48
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Title : SW846 8260
Last Update : Tue Jun 14 11:34:08 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.1	12754	PASS
75	95	30	60	59.6	31523	PASS
95	95	100	100	100.0	52859	PASS
96	95	5	9	7.5	3945	PASS
173	174	0.00	2	1.5	498	PASS
174	95	50	100	63.7	33661	PASS
175	174	5	9	8.2	2751	PASS
176	174	95	101	95.5	32136	PASS
177	176	5	9	7.0	2238	PASS