


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061322\
Data File : VN072804.D
Acq On : 13 Jun 2022 17:55
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061322W.M
Title : SW846 8260
Last Update : Tue Jun 14 11:34:08 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.1	9777	PASS
75	95	30	60	60.0	23408	PASS
95	95	100	100	100.0	39000	PASS
96	95	5	9	6.4	2477	PASS
173	174	0.00	2	1.1	270	PASS
174	95	50	100	60.4	23571	PASS
175	174	5	9	7.6	1800	PASS
176	174	95	101	97.4	22947	PASS
177	176	5	9	6.6	1513	PASS