

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
 Data File : VN078177.D
 Acq On : 13 Jun 2023 10:30
 Operator : JC\MD
 Sample : VN0613MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0613MBL01

Quant Time: Jun 14 02:15:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	466585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	863332	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	748191	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	230867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	332202	53.400	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.800%
35) Dibromofluoromethane	8.177	113	293844	53.830	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.660%
50) Toluene-d8	10.570	98	1003867	47.703	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.400%
62) 4-Bromofluorobenzene	12.853	95	307025	46.931	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
Data File : VN078177.D
Acq On : 13 Jun 2023 10:30
Operator : JC\MD
Sample : VN0613MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0613MBL01

Quant Time: Jun 14 02:15:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

