

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
 Data File : VN078183.D
 Acq On : 13 Jun 2023 14:00
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 14 02:19:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	423749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	766762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	679327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	216807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	273438	48.397	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.800%
35) Dibromofluoromethane	8.177	113	243092	50.141	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.280%
50) Toluene-d8	10.571	98	946555	50.645	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.280%
62) 4-Bromofluorobenzene	12.853	95	283666	48.821	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.640%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	11271	6.931	ug/l #	86
52) Toluene	10.635	92	13634	1.040	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
Data File : VN078183.D
Acq On : 13 Jun 2023 14:00
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 14 02:19:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

