

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
 Data File : VN078186.D
 Acq On : 13 Jun 2023 15:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 14 02:20:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	382265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	706002	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	602132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	180800	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	264249	51.846	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.700%	
35) Dibromofluoromethane	8.177	113	221618	49.646	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.300%	
50) Toluene-d8	10.570	98	861488	50.060	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.120%	
62) 4-Bromofluorobenzene	12.853	95	248606	46.470	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.940%	
Target Compounds						
16) Acetone	4.453	43	4080m	2.781	ug/l	Qvalue
52) Toluene	10.635	92	3069	0.254	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
Data File : VN078186.D
Acq On : 13 Jun 2023 15:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 14 02:20:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

