

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
 Data File : VN078178.D
 Acq On : 13 Jun 2023 10:54
 Operator : JC\MD
 Sample : VN0613WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0613WBL01

Quant Time: Jun 14 02:16:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	501772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	915589	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	814526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	268530	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	349881	52.298	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.600%
35) Dibromofluoromethane	8.171	113	311889	53.875	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.740%
50) Toluene-d8	10.571	98	1083810	48.563	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.120%
62) 4-Bromofluorobenzene	12.853	95	347548	50.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061323\
Data File : VN078178.D
Acq On : 13 Jun 2023 10:54
Operator : JC\MD
Sample : VN0613WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0613WBL01

Quant Time: Jun 14 02:16:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. 0.000 min
Lab File: VN078178.D
Acq: 13 Jun 2023 10:54




Instrument :
MSVOA_N
ClientSampleId :
VN0613WBL01


Tgt Ion:168 Resp: 501772
Ion Ratio Lower Upper
168 100
99 61.4 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 52.298 ug/l
RT: 8.582 min Scan# 1127
Delta R.T. -0.000 min
Lab File: VN078178.D
Acq: 13 Jun 2023 10:54

Tgt Ion: 65 Resp: 349881
Ion Ratio Lower Upper
65 100
67 52.9 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN078178.D

Acq: 13 Jun 2023 10:54

Instrument :

MSVOA_N

ClientSampleId :

VN0613WBL01





Tgt Ion:114 Resp: 915589

Ion Ratio Lower Upper

114 100

63 19.6 0.0 36.4

88 15.9 0.0 30.6



#35

Dibromofluoromethane

Concen: 53.875 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.006 min

Lab File: VN078178.D

Acq: 13 Jun 2023 10:54




Tgt Ion:113 Resp: 311889


Ion Ratio Lower Upper

113 100

111 104.6 82.6 123.8

192 16.9 12.7 19.1









#50
Toluene-d8
Concen: 48.563 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN078178.D
Acq: 13 Jun 2023 10:54




Instrument :
MSVOA_N
ClientSampleId :
VN0613WBL01


Tgt Ion: 98 Resp: 1083810
Ion Ratio Lower Upper
98 100
100 63.3 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 50.093 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VN078178.D
Acq: 13 Jun 2023 10:54

Tgt Ion: 95 Resp: 347548
Ion Ratio Lower Upper
95 100
174 74.8 0.0 143.8
176 69.7 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.870 min Scan# 1170
Delta R.T. 0.000 min
Lab File: VN078178.D
Acq: 13 Jun 2023 10:54

Instrument :
MSVOA_N
ClientSampleId :
VN0613WBL01

Tgt Ion:117 Resp: 814526
Ion Ratio Lower Upper
117 100
82 53.9 44.0 66.0
119 33.0 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.800 min Scan# 2014
Delta R.T. 0.000 min
Lab File: VN078178.D
Acq: 13 Jun 2023 10:54

Tgt Ion:152 Resp: 268530
Ion Ratio Lower Upper
152 100
115 61.6 32.3 96.8
150 156.6 0.0 355.8

