

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082570.D
 Acq On : 13 Jun 2024 19:35
 Operator : JC\MD
 Sample : P2838-08MSD
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-340-342MSD

Quant Time: Jun 14 02:35:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	152316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262175	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	244874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126444	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	125639	52.238	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.480%	
35) Dibromofluoromethane	8.171	113	91746	56.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 113.780%	
50) Toluene-d8	10.571	98	314621	50.883	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.760%	
62) 4-Bromofluorobenzene	12.853	95	123624	56.016	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 112.040%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	127258	57.949	ug/l	98
3) Chloromethane	2.359	50	142450	57.238	ug/l	97
4) Vinyl Chloride	2.512	62	141662	60.547	ug/l	99
5) Bromomethane	2.942	94	79619	59.204	ug/l	100
6) Chloroethane	3.106	64	97653	63.920	ug/l	99
7) Trichlorofluoromethane	3.489	101	184177	60.115	ug/l	97
8) Diethyl Ether	3.959	74	67164	61.537	ug/l	70
9) 1,1,2-Trichlorotrifluo...	4.365	101	103281	57.708	ug/l	98
10) Methyl Iodide	4.583	142	114015	63.370	ug/l	98
11) Tert butyl alcohol	5.524	59	132942	344.344	ug/l	99
12) 1,1-Dichloroethene	4.336	96	99484	59.805	ug/l #	79
13) Acrolein	4.177	56	92466	323.201	ug/l	96
14) Allyl chloride	5.018	41	193240	54.258	ug/l #	91
15) Acrylonitrile	5.724	53	340229	340.269	ug/l	99
16) Acetone	4.430	43	312889	317.133	ug/l	91
17) Carbon Disulfide	4.712	76	285559	53.150	ug/l	99
18) Methyl Acetate	5.024	43	161073	65.555	ug/l #	86
19) Methyl tert-butyl Ether	5.800	73	382625	62.928	ug/l	94
20) Methylene Chloride	5.277	84	118902	59.731	ug/l #	77
21) trans-1,2-Dichloroethene	5.789	96	106931	58.988	ug/l #	76
22) Diisopropyl ether	6.671	45	454758	61.149	ug/l #	93
23) Vinyl Acetate	6.606	43	1740442	296.830	ug/l #	90
24) 1,1-Dichloroethane	6.571	63	231683	60.441	ug/l	96
25) 2-Butanone	7.489	43	498746	333.415	ug/l #	85
26) 2,2-Dichloropropane	7.494	77	162815	50.711	ug/l	98
27) cis-1,2-Dichloroethene	7.489	96	132047	61.723	ug/l	84
28) Bromochloromethane	7.818	49	83343	50.028	ug/l #	76
29) Tetrahydrofuran	7.841	42	341036	355.086	ug/l #	82
30) Chloroform	7.965	83	226872	61.979	ug/l	96
31) Cyclohexane	8.259	56	199389	51.906	ug/l	87
32) 1,1,1-Trichloroethane	8.171	97	204632	63.013	ug/l	94
36) 1,1-Dichloropropene	8.371	75	157659	58.287	ug/l	97
37) Ethyl Acetate	7.565	43	195556	64.602	ug/l #	92
38) Carbon Tetrachloride	8.365	117	176720	64.450	ug/l	98
39) Methylcyclohexane	9.606	83	163119	55.017	ug/l #	87
40) Benzene	8.606	78	492806	61.097	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082570.D
 Acq On : 13 Jun 2024 19:35
 Operator : JC\MD
 Sample : P2838-08MSD
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-340-342MSD

Quant Time: Jun 14 02:35:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	111804	62.802	ug/l #	86
42) 1,2-Dichloroethane	8.671	62	190606	61.442	ug/l	96
43) Isopropyl Acetate	8.694	43	396183	73.862	ug/l #	84
44) Trichloroethene	9.353	130	107226	58.522	ug/l	98
45) 1,2-Dichloropropane	9.624	63	130915	62.906	ug/l	99
46) Dibromomethane	9.712	93	86924	63.619	ug/l	98
47) Bromodichloromethane	9.888	83	186266	64.576	ug/l #	95
48) Methyl methacrylate	9.683	41	164970	64.546	ug/l #	83
49) 1,4-Dioxane	9.694	88	56504	1577.681	ug/l #	80
51) 4-Methyl-2-Pentanone	10.447	43	1081629	354.918	ug/l	90
52) Toluene	10.635	92	306349	63.289	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	191145	64.075	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	192155	60.853	ug/l #	80
55) 1,1,2-Trichloroethane	11.018	97	115630	65.734	ug/l	99
56) Ethyl methacrylate	10.877	69	203465	68.128	ug/l #	74
57) 1,3-Dichloropropane	11.165	76	207507	63.092	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.312	63	1288	0.836	ug/l #	44
59) 2-Hexanone	11.200	43	819399	367.207	ug/l	90
60) Dibromochloromethane	11.359	129	140232	72.239	ug/l	99
61) 1,2-Dibromoethane	11.471	107	118760	65.730	ug/l	98
64) Tetrachloroethene	11.106	164	94044	51.815	ug/l	95
65) Chlorobenzene	11.894	112	323993	59.892	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.965	131	118785	66.662	ug/l	100
67) Ethyl Benzene	11.965	91	588651	60.829	ug/l	95
68) m/p-Xylenes	12.071	106	445741	125.678	ug/l	94
69) o-Xylene	12.400	106	213391	62.112	ug/l	92
70) Styrene	12.412	104	361649	63.374	ug/l	96
71) Bromoform	12.576	173	96113	67.568	ug/l #	98
73) Isopropylbenzene	12.700	105	547068	54.460	ug/l	99
74) N-amyl acetate	12.494	43	304713	55.147	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.941	83	177270	61.219	ug/l	97
76) 1,2,3-Trichloropropane	12.994	75	242089	90.965	ug/l	61
77) Bromobenzene	12.982	156	135021	56.402	ug/l	88
78) n-propylbenzene	13.035	91	656821	55.217	ug/l	96
79) 2-Chlorotoluene	13.129	91	401779	53.789	ug/l	94
80) 1,3,5-Trimethylbenzene	13.176	105	458807	56.485	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.741	75	74480	65.458	ug/l	90
82) 4-Chlorotoluene	13.223	91	405110	55.471	ug/l	95
83) tert-Butylbenzene	13.441	119	388682	54.698	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	462977	55.713	ug/l	98
85) sec-Butylbenzene	13.618	105	541139	54.941	ug/l	98
86) p-Isopropyltoluene	13.729	119	454866	56.825	ug/l	98
87) 1,3-Dichlorobenzene	13.735	146	245536	56.613	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	244289	55.188	ug/l	98
89) n-Butylbenzene	14.059	91	382155	53.116	ug/l	98
90) Hexachloroethane	14.335	117	88931	59.963	ug/l	97
91) 1,2-Dichlorobenzene	14.106	146	234437	57.106	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	38026	63.915	ug/l	89
93) 1,2,4-Trichlorobenzene	15.394	180	120988	52.576	ug/l	98
94) Hexachlorobutadiene	15.500	225	50078	49.594	ug/l	99
95) Naphthalene	15.641	128	419729	58.202	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	121742	52.847	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082570.D
Acq On : 13 Jun 2024 19:35
Operator : JC\MD
Sample : P2838-08MSD
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD

Quant Time: Jun 14 02:35:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.230 min Scan# 10

Delta R.T. 0.006 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-196-GW-340-342MSD

Tgt Ion:168 Resp: 152316

Ion Ratio Lower Upper

168 100

99 62.1 48.2 72.4



#2

Dichlorodifluoromethane

Concen: 57.949 ug/l

RT: 2.124 min Scan# 29

Delta R.T. 0.000 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion: 85 Resp: 127258

Ion Ratio Lower Upper

85 100

87 32.8 15.8 47.4





#3

Chloromethane

Concen: 57.238 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-196-GW-340-342MSD

Tgt Ion: 50 Resp: 142450

Ion Ratio Lower Upper

50 100

52 33.3 25.5 38.3



#4

Vinyl Chloride

Concen: 60.547 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion: 62 Resp: 141662

Ion Ratio Lower Upper

62 100

64 30.1 24.6 37.0









#5
Bromomethane
Concen: 59.204 ug/l
RT: 2.942 min Scan# 1
Delta R.T. -0.012 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 94 Resp: 79619
Ion Ratio Lower Upper
94 100
96 93.7 74.8 112.2



#6
Chloroethane
Concen: 63.920 ug/l
RT: 3.106 min Scan# 196
Delta R.T. -0.012 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 64 Resp: 97653
Ion Ratio Lower Upper
64 100
66 30.6 24.7 37.1









#7
Trichlorofluoromethane
Concen: 60.115 ug/l
RT: 3.489 min Scan# 201
Delta R.T. -0.012 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:101 Resp: 184177
Ion Ratio Lower Upper
101 100
103 65.5 50.6 75.8



#8
Diethyl Ether
Concen: 61.537 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 74 Resp: 67164
Ion Ratio Lower Upper
74 100
45 112.7 42.4 127.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 57.708 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:101 Resp: 103281
Ion Ratio Lower Upper
101 100
85 45.8 37.3 55.9
151 71.6 59.2 88.8



#10
Methyl Iodide
Concen: 63.370 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:142 Resp: 114015
Ion Ratio Lower Upper
142 100
127 46.5 35.8 53.6
141 14.3 11.4 17.2





#11

Tert butyl alcohol

Concen: 344.344 ug/l

RT: 5.524 min Scan# 606

Delta R.T. 0.006 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-196-GW-340-342MSD

Tgt Ion: 59 Resp: 132942

Ion Ratio Lower Upper

59 100

57 10.9 8.9 13.3



#12

1,1-Dichloroethene

Concen: 59.805 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion: 96 Resp: 99484


Ion Ratio Lower Upper

96 100

61 186.4 119.7 179.5#

98 62.7 49.6 74.4









#13
Acrolein
Concen: 323.201 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 56 Resp: 92466
Ion Ratio Lower Upper
56 100
55 73.4 55.8 83.8



#14
Allyl chloride
Concen: 54.258 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 41 Resp: 193240
Ion Ratio Lower Upper
41 100
39 83.0 65.4 98.0
76 32.3 37.3 55.9#





#15

Acrylonitrile

Concen: 340.269 ug/l

RT: 5.724 min Scan# 641

Delta R.T. 0.006 min

Lab File: VN082570.D


Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD





Tgt Ion: 53 Resp: 340229

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

53	100		
----	-----	--	--

52	83.1	66.1	99.1
----	------	------	------

51	37.8	29.4	44.2
----	------	------	------



#16

Acetone

Concen: 317.133 ug/l

RT: 4.430 min Scan# 421

Delta R.T. 0.006 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion: 43 Resp: 312889

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

43	100		
----	-----	--	--

58	26.6	25.2	37.8
----	------	------	------









#17
Carbon Disulfide
Concen: 53.150 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 76 Resp: 285559
Ion Ratio Lower Upper
76 100
78 8.9 7.5 11.3



#18
Methyl Acetate
Concen: 65.555 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 43 Resp: 161073
Ion Ratio Lower Upper
43 100
74 20.3 22.2 33.2#









#19
Methyl tert-butyl Ether
Concen: 62.928 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 73 Resp: 382625
Ion Ratio Lower Upper
73 100
57 23.2 16.5 24.7



#20
Methylene Chloride
Concen: 59.731 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 84 Resp: 118902
Ion Ratio Lower Upper
84 100
49 146.9 84.9 127.3#
51 44.0 27.0 40.6#
86 65.5 52.3 78.5









#21
trans-1,2-Dichloroethene
Concen: 58.988 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 96 Resp: 106931
Ion Ratio Lower Upper
96 100
61 160.5 96.5 144.7#
98 61.8 51.3 76.9



#22
Diisopropyl ether
Concen: 61.149 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 45 Resp: 454758
Ion Ratio Lower Upper
45 100
43 56.2 44.1 66.1
87 23.6 27.0 40.6#
59 10.6 9.6 14.4









#23
 Vinyl Acetate
 Concen: 296.830 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35




Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-340-342MSD


Tgt Ion: 43 Resp: 1740442
 Ion Ratio Lower Upper
 43 100
 86 8.7 10.2 15.4#



#24
 1,1-Dichloroethane
 Concen: 60.441 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35

Tgt Ion: 63 Resp: 231683
 Ion Ratio Lower Upper
 63 100
 98 6.0 3.6 10.9
 100 4.1 2.6 7.8









#25
2-Butanone
Concen: 333.415 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 43 Resp: 498746
Ion Ratio Lower Upper
43 100
72 23.0 25.3 37.9#



#26
2,2-Dichloropropane
Concen: 50.711 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 77 Resp: 162815
Ion Ratio Lower Upper
77 100
97 21.6 11.3 33.9









#27
 cis-1,2-Dichloroethene
 Concen: 61.723 ug/l
 RT: 7.489 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35




Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-340-342MSD


Tgt Ion: 96 Resp: 132047
 Ion Ratio Lower Upper
 96 100
 61 161.9 0.0 268.0
 98 63.5 0.0 126.8



#28
 Bromochloromethane
 Concen: 50.028 ug/l
 RT: 7.818 min Scan# 997
 Delta R.T. 0.006 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35

Tgt Ion: 49 Resp: 83343
 Ion Ratio Lower Upper
 49 100
 129 1.1 0.0 4.6
 130 61.2 66.7 100.1#





#29

Tetrahydrofuran

Concen: 355.086 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.005 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD





Tgt Ion: 42 Resp: 341036

Ion Ratio Lower Upper

42 100

72 36.8 39.3 58.9#

71 35.2 37.4 56.0#



#30

Chloroform

Concen: 61.979 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion: 83 Resp: 226872

Ion Ratio Lower Upper

83 100

85 66.8 50.7 76.1









#31
Cyclohexane
Concen: 51.906 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 56 Resp: 199389
Ion Ratio Lower Upper
56 100
69 30.0 24.4 36.6
84 79.8 77.0 115.4



#32
1,1,1-Trichloroethane
Concen: 63.013 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 97 Resp: 204632
Ion Ratio Lower Upper
97 100
99 64.8 50.3 75.5
61 52.7 36.6 55.0





#33

1,2-Dichloroethane-d4

Concen: 52.238 ug/l

RT: 8.583 min Scan# 1127

Delta R.T. 0.006 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-196-GW-340-342MSD

Tgt Ion: 65 Resp: 125639

Ion Ratio Lower Upper

65 100

67 50.4 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.006 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion: 114 Resp: 262175


Ion Ratio Lower Upper

114 100

63 23.5 0.0 41.0

88 15.8 0.0 31.4









#35
 Dibromofluoromethane
 Concen: 56.887 ug/l
 RT: 8.171 min Scan# 1056
 Delta R.T. 0.006 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35




Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-340-342MSD


Tgt Ion:113 Resp: 91746
 Ion Ratio Lower Upper
 113 100
 111 103.5 83.4 125.2
 192 19.0 14.0 21.0



#36
 1,1-Dichloropropene
 Concen: 58.287 ug/l
 RT: 8.371 min Scan# 1091
 Delta R.T. 0.000 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35

Tgt Ion: 75 Resp: 157659
 Ion Ratio Lower Upper
 75 100
 110 32.5 17.8 53.5
 77 30.7 25.0 37.4





#37

Ethyl Acetate

Concen: 64.602 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN082570.D


Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD





Tgt Ion: 43 Resp: 195556

Ion Ratio Lower Upper

43 100

61 10.3 11.3 16.9#

70 8.4 8.7 13.1#



#38

Carbon Tetrachloride

Concen: 64.450 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion: 117 Resp: 176720


Ion Ratio Lower Upper

117 100

119 93.7 76.5 114.7

121 30.8 23.9 35.9









#39
Methylcyclohexane
Concen: 55.017 ug/l
RT: 9.606 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 83 Resp: 163119
Ion Ratio Lower Upper
83 100
55 83.5 53.0 79.6#
98 46.8 37.1 55.7



#40
Benzene
Concen: 61.097 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 78 Resp: 492806
Ion Ratio Lower Upper
78 100
77 24.4 20.0 30.0









#41
Methacrylonitrile
Concen: 62.802 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:	41	Resp:	111804
Ion	Ratio	Lower	Upper
41	100		
39	61.0	46.6	69.8
67	62.4	67.3	100.9
52	30.0	27.3	40.9



#42
1,2-Dichloroethane
Concen: 61.442 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35


Tgt Ion:	62	Resp:	190606
Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	19.6








#43
Isopropyl Acetate
Concen: 73.862 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD







Tgt Ion: 43 Resp: 396183
Ion Ratio Lower Upper
43 100
61 19.6 22.8 34.2#
87 9.2 12.0 18.0#



#44
Trichloroethene
Concen: 58.522 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:130 Resp: 107226
Ion Ratio Lower Upper
130 100
95 99.8 0.0 195.4





#45

1,2-Dichloropropane

Concen: 62.906 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-196-GW-340-342MSD

Tgt Ion: 63 Resp: 130915

Ion Ratio Lower Upper

63 100

65 29.6 24.1 36.1



#46

Dibromomethane

Concen: 63.619 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion: 93 Resp: 86924


Ion Ratio Lower Upper

93 100

95 79.9 64.8 97.2

174 89.9 73.4 110.0





#47

Bromodichloromethane

Concen: 64.576 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN082570.D


Acq: 13 Jun 2024 19:35

Instrument :




MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD



Tgt Ion:	83	Resp:	186266
Ion Ratio	Lower	Upper	
83	100		
85	61.4	51.9	77.9
127	7.3	7.5	11.3



#48

Methyl methacrylate

Concen: 64.546 ug/l




RT: 9.683 min Scan# 1314


Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Tgt Ion:	41	Resp:	164970
Ion Ratio	Lower	Upper	
41	100		
69	70.1	77.9	116.9
39	62.9	50.9	76.3





#49

1,4-Dioxane

Concen: 1577.681 ug/l

RT: 9.694 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN082570.D


Acq: 13 Jun 2024 19:35

Instrument :




MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD



Tgt Ion:	88	Resp:	56504
Ion	Ratio	Lower	Upper
88	100		
43	37.9	19.0	28.6#
58	76.2	50.6	76.0#



#50

Toluene-d8

Concen: 50.883 ug/l




RT: 10.571 min Scan# 1465


Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Tgt Ion:	98	Resp:	314621
Ion	Ratio	Lower	Upper
98	100		
100	64.9	51.8	77.6









#51
4-Methyl-2-Pentanone
Concen: 354.918 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 43 Resp: 1081629
Ion Ratio Lower Upper
43 100
58 35.0 33.2 49.8



#52
Toluene
Concen: 63.289 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 92 Resp: 306349
Ion Ratio Lower Upper
92 100
91 173.7 136.7 205.1









#53
t-1,3-Dichloropropene
Concen: 64.075 ug/l
RT: 10.835 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 75 Resp: 191145
Ion Ratio Lower Upper
75 100
77 31.0 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 60.853 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 75 Resp: 192155
Ion Ratio Lower Upper
75 100
77 30.0 25.0 37.4
39 65.0 34.8 52.2#





#55

1,1,2-Trichloroethane

Concen: 65.734 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.000 min

Lab File: VN082570.D


Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD



Tgt Ion: 97 Resp: 115630




Ion Ratio Lower Upper

97 100

83 88.3 70.4 105.6

85 58.1 45.6 68.4

99 62.4 50.9 76.3



#56

Ethyl methacrylate

Concen: 68.128 ug/l

RT: 10.877 min Scan# 1517

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion: 69 Resp: 203465


Ion Ratio Lower Upper

69 100

41 78.8 46.6 70.0#

39 45.5 26.5 39.7#





#57

1,3-Dichloropropane

Concen: 63.092 ug/l

RT: 11.165 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-196-GW-340-342MSD

Tgt Ion: 76 Resp: 207507

Ion Ratio Lower Upper

76 100

78 31.7 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 0.836 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.153 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion: 63 Resp: 1288

Ion Ratio Lower Upper

63 100

106 0.0 24.6 36.8#









#59
2-Hexanone
Concen: 367.207 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 43 Resp: 819399
Ion Ratio Lower Upper
43 100
58 49.0 28.3 84.9



#60
Dibromochloromethane
Concen: 72.239 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 129 Resp: 140232
Ion Ratio Lower Upper
129 100
127 77.1 38.3 114.8









#61
1,2-Dibromoethane
Concen: 65.730 ug/l
RT: 11.471 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:107 Resp: 118760
Ion Ratio Lower Upper
107 100
109 93.0 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 56.016 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 95 Resp: 123624
Ion Ratio Lower Upper
95 100
174 77.3 0.0 160.6
176 73.7 0.0 152.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:117 Resp: 244874
Ion Ratio Lower Upper
117 100
82 59.7 44.6 66.8
119 31.4 25.9 38.9



#64
Tetrachloroethene
Concen: 51.815 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:164 Resp: 94044
Ion Ratio Lower Upper
164 100
166 128.7 97.6 146.4
129 97.2 73.7 110.5
131 90.6 75.8 113.6





#65

Chlorobenzene

Concen: 59.892 ug/l

RT: 11.894 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :





BP-VPB-196-GW-340-342MSD

Tgt Ion:112 Resp: 323993

Ion Ratio Lower Upper

112 100

114 32.4 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 66.662 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.006 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion:131 Resp: 118785


Ion Ratio Lower Upper

131 100

133 95.1 47.5 142.6

119 65.5 32.9 98.7









#67
Ethyl Benzene
Concen: 60.829 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 91 Resp: 588651
Ion Ratio Lower Upper
91 100
106 29.0 25.3 37.9



#68
m/p-Xylenes
Concen: 125.678 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:106 Resp: 445741
Ion Ratio Lower Upper
106 100
91 207.6 159.4 239.0









#69
o-Xylene
Concen: 62.112 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:106 Resp: 213391
Ion Ratio Lower Upper
106 100
91 221.1 104.7 314.1



#70
Styrene
Concen: 63.374 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:104 Resp: 361649
Ion Ratio Lower Upper
104 100
78 53.4 38.6 58.0
103 55.7 43.8 65.6





#71

Bromoform

Concen: 67.568 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD





Tgt Ion:173 Resp: 96113

Ion Ratio Lower Upper

173 100

175 47.8 24.6 74.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion:152 Resp: 126444


Ion Ratio Lower Upper

152 100

115 57.9 30.4 91.3

150 172.4 0.0 353.2









#73
Isopropylbenzene
Concen: 54.460 ug/l
RT: 12.700 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:105 Resp: 547068
Ion Ratio Lower Upper
105 100
120 26.2 13.4 40.1



#74
N-ethyl acetate
Concen: 55.147 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 43 Resp: 304713
Ion Ratio Lower Upper
43 100
70 37.1 39.0 58.4#
55 22.8 21.3 31.9
61 21.3 20.5 30.7





#75

1,1,2,2-Tetrachloroethane

Concen: 61.219 ug/l

RT: 12.941 min Scan# 1868

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD





Tgt Ion: 83 Resp: 177270

Ion Ratio Lower Upper

83 100

131 9.8 5.5 16.4

85 63.5 32.9 98.6



#76

1,2,3-Trichloropropane

Concen: 90.965 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.000 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion: 75 Resp: 242089

Ion Ratio Lower Upper

75 100

77 126.1 90.5 271.3









#77
Bromobenzene
Concen: 56.402 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:156 Resp: 135021
Ion Ratio Lower Upper
156 100
77 226.1 100.2 300.6
158 95.2 49.1 147.3



#78
n-propylbenzene
Concen: 55.217 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 91 Resp: 656821
Ion Ratio Lower Upper
91 100
120 21.4 11.7 35.0









#79
2-Chlorotoluene
Concen: 53.789 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 91 Resp: 401779
Ion Ratio Lower Upper
91 100
126 31.6 17.5 52.5



#80
1,3,5-Trimethylbenzene
Concen: 56.485 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 105 Resp: 458807
Ion Ratio Lower Upper
105 100
120 47.9 24.8 74.3









#81
trans-1,4-Dichloro-2-butene
Concen: 65.458 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion: 75 Resp: 74480
Ion Ratio Lower Upper
75 100
53 99.1 72.9 109.3
89 40.6 39.1 58.7



#82
4-Chlorotoluene
Concen: 55.471 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion: 91 Resp: 405110
Ion Ratio Lower Upper
91 100
126 32.1 17.4 52.3





#83

tert-Butylbenzene

Concen: 54.698 ug/l

RT: 13.441 min Scan# 1953

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD





Tgt Ion:119 Resp: 388682

Ion Ratio Lower Upper

119 100

91 69.8 32.6 97.8

134 24.6 12.4 37.0



#84

1,2,4-Trimethylbenzene

Concen: 55.713 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion:105 Resp: 462977

Ion Ratio Lower Upper

105 100

120 44.6 22.9 68.7









#85
 sec-Butylbenzene
 Concen: 54.941 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35




Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-340-342MSD


Tgt Ion:105 Resp: 541139
 Ion Ratio Lower Upper
 105 100
 134 19.5 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 56.825 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN082570.D
 Acq: 13 Jun 2024 19:35

Tgt Ion:119 Resp: 454866
 Ion Ratio Lower Upper
 119 100
 134 25.1 13.2 39.5
 91 24.4 11.8 35.3









#87
1,3-Dichlorobenzene
Concen: 56.613 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:146 Resp: 245536
Ion Ratio Lower Upper
146 100
111 41.7 20.8 62.5
148 63.1 32.1 96.3



#88
1,4-Dichlorobenzene
Concen: 55.188 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:146 Resp: 244289
Ion Ratio Lower Upper
146 100
111 39.8 19.3 57.9
148 63.8 31.3 93.9





#89

n-Butylbenzene

Concen: 53.116 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD





Tgt Ion: 91 Resp: 382155

Ion Ratio Lower Upper

91 100

92 51.7 26.2 78.5

134 24.2 13.2 39.5



#90

Hexachloroethane

Concen: 59.963 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN082570.D




Acq: 13 Jun 2024 19:35


Tgt Ion: 117 Resp: 88931

Ion Ratio Lower Upper

117 100

201 68.9 33.3 99.8









#91
1,2-Dichlorobenzene
Concen: 57.106 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD


Tgt Ion:	146	Resp:	234437
Ion	Ratio	Lower	Upper
146	100		
111	42.8	21.3	64.0
148	64.6	31.8	95.4



#92
1,2-Dibromo-3-Chloropropane
Concen: 63.915 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.006 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:	75	Resp:	38026
Ion	Ratio	Lower	Upper
75	100		
155	73.9	42.6	128.0
157	95.2	53.0	159.0





#93

1,2,4-Trichlorobenzene

Concen: 52.576 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN082570.D


Acq: 13 Jun 2024 19:35

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-196-GW-340-342MSD






Tgt Ion:180 Resp: 120988

Ion Ratio Lower Upper

180 100

182 94.2 46.7 140.0

145 30.1 16.2 48.6



#94

Hexachlorobutadiene

Concen: 49.594 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN082570.D

Acq: 13 Jun 2024 19:35




Tgt Ion:225 Resp: 50078


Ion Ratio Lower Upper

225 100

223 61.9 31.1 93.2

227 62.8 30.6 91.6









#95
Naphthalene
Concen: 58.202 ug/l
RT: 15.641 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342MSD

Tgt Ion:	128	Resp:	419729
Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.1	15.1
129	10.9	8.9	13.3



#96
1,2,3-Trichlorobenzene
Concen: 52.847 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN082570.D
Acq: 13 Jun 2024 19:35

Tgt Ion:	180	Resp:	121742
Ion	Ratio	Lower	Upper
180	100		
182	95.5	46.0	137.8
145	33.4	17.1	51.3

