

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082548.D
 Acq On : 13 Jun 2024 10:31
 Operator : JC\MD
 Sample : VN0613MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0613MBL01

Quant Time: Jun 14 02:45:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	328203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	326940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	137946	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	154003	55.056	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.120%
35) Dibromofluoromethane	8.165	113	111216	55.086	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.180%
50) Toluene-d8	10.571	98	423755	54.745	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.500%
62) 4-Bromofluorobenzene	12.853	95	145682	52.730	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	105.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082548.D
Acq On : 13 Jun 2024 10:31
Operator : JC\MD
Sample : VN0613MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0613MBL01

Quant Time: Jun 14 02:45:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

