

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082559.D
 Acq On : 13 Jun 2024 15:07
 Operator : JC\MD
 Sample : P2853-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 7

Quant Time: Jun 14 02:29:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	302533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	304227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	132593	53.028	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.060%	
35) Dibromofluoromethane	8.171	113	95827	51.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.980%	
50) Toluene-d8	10.571	98	374867	52.538	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	12.853	95	123582	48.527	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 97.060%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	21550	21.009	ug/l	95
20) Methylene Chloride	5.277	84	12348	5.967	ug/l #	73
25) 2-Butanone	7.494	43	11162	7.177	ug/l #	85
43) Isopropyl Acetate	8.694	43	279418	45.144	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082559.D
Acq On : 13 Jun 2024 15:07
Operator : JC\MD
Sample : P2853-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
7

Quant Time: Jun 14 02:29:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

