

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082561.D
 Acq On : 13 Jun 2024 15:56
 Operator : JC\MD
 Sample : P2853-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 9

Quant Time: Jun 14 02:30:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	163430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	311768	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	316623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127391	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	136391	52.852	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.700%	
35) Dibromofluoromethane	8.171	113	101177	52.755	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.520%	
50) Toluene-d8	10.571	98	385360	52.409	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.820%	
62) 4-Bromofluorobenzene	12.853	95	127085	48.424	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 96.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	21415	20.229	ug/l #	85
20) Methylene Chloride	5.271	84	12372	5.793	ug/l #	73
25) 2-Butanone	7.488	43	10855	6.763	ug/l #	87
43) Isopropyl Acetate	8.694	43	290466	45.539	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082561.D
Acq On : 13 Jun 2024 15:56
Operator : JC\MD
Sample : P2853-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
9

Quant Time: Jun 14 02:30:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

