

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082563.D
 Acq On : 13 Jun 2024 16:44
 Operator : JC\MD
 Sample : P2853-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11

Quant Time: Jun 14 02:31:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161595	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	308454	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127565	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	136183	53.371	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.740%	
35) Dibromofluoromethane	8.171	113	98738	52.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.080%	
50) Toluene-d8	10.571	98	385210	52.952	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.900%	
62) 4-Bromofluorobenzene	12.853	95	124788	48.060	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 96.120%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	21812	20.838	ug/l	94
20) Methylene Chloride	5.277	84	12297	5.823	ug/l #	81
25) 2-Butanone	7.488	43	11823	7.450	ug/l #	75
37) Ethyl Acetate	7.565	43	4963	1.394	ug/l #	80
43) Isopropyl Acetate	8.694	43	282811	44.815	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082563.D
Acq On : 13 Jun 2024 16:44
Operator : JC\MD
Sample : P2853-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11

Quant Time: Jun 14 02:31:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

