

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082564.D
 Acq On : 13 Jun 2024 17:09
 Operator : JC\MD
 Sample : P2853-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 12

Quant Time: Jun 14 02:32:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	296170	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	309375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	132065	54.400	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.800%
35) Dibromofluoromethane	8.171	113	96469	52.950	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.900%
50) Toluene-d8	10.570	98	382440	54.751	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.500%
62) 4-Bromofluorobenzene	12.853	95	125784	50.452	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.900%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	21751	21.841	ug/l	98
20) Methylene Chloride	5.283	84	12993	6.466	ug/l #	73
25) 2-Butanone	7.488	43	12058	7.986	ug/l #	85
43) Isopropyl Acetate	8.688	43	296286	48.898	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082564.D
Acq On : 13 Jun 2024 17:09
Operator : JC\MD
Sample : P2853-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
12

Quant Time: Jun 14 02:32:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

