

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082566.D
 Acq On : 13 Jun 2024 17:58
 Operator : JC\MD
 Sample : P2864-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-H

Quant Time: Jun 14 02:33:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	161826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	316975	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	321121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	130601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	136289	53.336	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.680%	
35) Dibromofluoromethane	8.171	113	101394	52.000	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.000%	
50) Toluene-d8	10.570	98	390149	52.189	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.380%	
62) 4-Bromofluorobenzene	12.853	95	127519	47.791	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 95.580%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	25854	24.665	ug/l	89
20) Methylene Chloride	5.277	84	7826	3.700	ug/l #	70
25) 2-Butanone	7.494	43	13057	8.216	ug/l #	78
43) Isopropyl Acetate	8.694	43	278614	42.963	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082566.D
Acq On : 13 Jun 2024 17:58
Operator : JC\MD
Sample : P2864-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-H

Quant Time: Jun 14 02:33:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

