

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
 Data File : VN082567.D
 Acq On : 13 Jun 2024 18:22
 Operator : JC\MD
 Sample : P2864-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-I

Quant Time: Jun 14 02:33:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	163974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	316572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	320999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	138795	53.606	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.220%	
35) Dibromofluoromethane	8.171	113	101473	52.107	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.220%	
50) Toluene-d8	10.571	98	399360	53.489	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.980%	
62) 4-Bromofluorobenzene	12.853	95	132497	49.720	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 99.440%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	20797	19.580	ug/l	91
20) Methylene Chloride	5.277	84	8162	3.809	ug/l #	79
25) 2-Butanone	7.489	43	11838	7.351	ug/l	93
43) Isopropyl Acetate	8.688	43	321464	49.634	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061324\
Data File : VN082567.D
Acq On : 13 Jun 2024 18:22
Operator : JC\MD
Sample : P2864-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-I

Quant Time: Jun 14 02:33:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

