

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061325\
 Data File : VN087002.D
 Acq On : 13 Jun 2025 15:28
 Operator : JC\MD
 Sample : Q2314-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VPB182-HYD-20250612

Quant Time: Jun 14 00:29:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	207879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	402501	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	359714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	173766	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	141723	50.920	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.840%		
35) Dibromofluoromethane	8.177	113	120918	50.693	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.380%		
50) Toluene-d8	10.571	98	485254	51.389	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.780%		
62) 4-Bromofluorobenzene	12.847	95	178616	50.913	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 101.820%		
Target Compounds						
60) Dibromochloromethane	11.365	129	3371	1.184	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061325\
Data File : VN087002.D
Acq On : 13 Jun 2025 15:28
Operator : JC\MD
Sample : Q2314-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VPB182-HYD-20250612

Quant Time: Jun 14 00:29:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

