

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061419\
 Data File : VN056301.D
 Acq On : 14 Jun 2019 9:50
 Operator : JC/SP
 Sample : VN0614WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0614WBL01

Quant Time: Jun 14 10:16:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	56542	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	286593	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	220737	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	114143	32.51	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	108.37%
60) 4-Bromofluorobenzene	12.40	95	65447	20.00	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	66.67%
63) Toluene-d8	10.09	98	330847	33.04	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	110.13%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061419\
Data File : VN056301.D
Acq On : 14 Jun 2019 9:50
Operator : JC/SP
Sample : VN0614WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0614WBL01

Quant Time: Jun 14 10:16:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jun 06 04:02:53 2019
Response via : Initial Calibration

