

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061419\
 Data File : VN056304.D
 Acq On : 14 Jun 2019 11:01
 Operator : JC/SP
 Sample : K3372-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFFL-B2-79

Quant Time: Jun 14 11:24:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	58523	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	310460	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	243377	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	122551	33.72	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	112.40%
60) 4-Bromofluorobenzene	12.40	95	81375	22.56	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	75.20%
63) Toluene-d8	10.09	98	369546	33.47	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	111.57%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061419\
Data File : VN056304.D
Acq On : 14 Jun 2019 11:01
Operator : JC/SP
Sample : K3372-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFFL-B2-79

Quant Time: Jun 14 11:24:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jun 06 04:02:53 2019
Response via : Initial Calibration

