

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061419\
 Data File : VN056306.D
 Acq On : 14 Jun 2019 11:48
 Operator : JC/SP
 Sample : K3338-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 55-ST-17

Quant Time: Jun 14 23:31:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	59530	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	308927	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	250418	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	123822	33.49	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	111.63%
60) 4-Bromofluorobenzene	12.40	95	82843	22.32	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.40%
63) Toluene-d8	10.09	98	369217	32.50	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	108.33%

Target Compounds

					Ovalue
15) Acetone	3.83	58	2679	4.913	ug/l 94
22) cis-1,2-Dichloroethene	6.83	96	14431	3.728	ug/l 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061419\
Data File : VN056306.D
Acq On : 14 Jun 2019 11:48
Operator : JC/SP
Sample : K3338-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
55-ST-17

Quant Time: Jun 14 23:31:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jun 06 04:02:53 2019
Response via : Initial Calibration

