

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061419\
 Data File : VN056308.D
 Acq On : 14 Jun 2019 13:40
 Operator : JC/SP
 Sample : VN0614WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0614WBL02

Quant Time: Jun 14 23:57:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	341509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	567913	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	540098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	150092	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	192647	56.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.62%	
35) Dibromofluoromethane	7.59	113	179383	51.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.96%	
50) Toluene-d8	10.09	98	697860	52.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.10%	
62) 4-Bromofluorobenzene	12.40	95	211847	46.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061419\
Data File : VN056308.D
Acq On : 14 Jun 2019 13:40
Operator : JC/SP
Sample : VN0614WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0614WBL02

Quant Time: Jun 14 23:57:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

