

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061419\
 Data File : VN056321.D
 Acq On : 14 Jun 2019 18:50
 Operator : JC/SP
 Sample : K3352-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190612-LOC-11

Quant Time: Jun 15 00:41:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	322826	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	543338	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	559955	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	176112	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	186789	58.27	ug/l	0.00
Spiked Amount						
			Recovery	=	116.54%	
35) Dibromofluoromethane	7.59	113	170201	51.06	ug/l	0.00
Spiked Amount						
			Recovery	=	102.12%	
50) Toluene-d8	10.09	98	691645	53.92	ug/l	0.00
Spiked Amount						
			Recovery	=	107.84%	
62) 4-Bromofluorobenzene	12.40	95	229947	52.71	ug/l	0.00
Spiked Amount						
			Recovery	=	105.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	42551	27.756	ug/l	97
18) Methyl Acetate	4.34	43	6531	1.609	ug/l #	80
20) Methylene Chloride	4.54	84	277461	87.848	ug/l	97
30) Chloroform	7.37	83	25061	4.882	ug/l	93
43) Isopropyl Acetate	8.17	43	11248	1.670	ug/l #	87
95) Naphthalene	15.13	128	10377	4.988	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061419\
Data File : VN056321.D
Acq On : 14 Jun 2019 18:50
Operator : JC/SP
Sample : K3352-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190612-LOC-11

Quant Time: Jun 15 00:41:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

