

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072843.D
 Acq On : 14 Jun 2022 19:47
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 15 04:35:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	350082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	580278	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	518147	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	207670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	224463	41.399	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	82.800%
35) Dibromofluoromethane	8.022	113	181193	50.434	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.860%
50) Toluene-d8	10.439	98	616254	43.465	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	86.920%
62) 4-Bromofluorobenzene	12.727	95	237103	49.818	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072843.D
Acq On : 14 Jun 2022 19:47
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 15 04:35:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

