

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072849.D
 Acq On : 14 Jun 2022 23:03
 Operator : JC\MD
 Sample : PB145494ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145494ZHE#01

Quant Time: Jun 15 04:36:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	327153	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	553295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	484862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	189449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	211300	41.702	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.400%
35) Dibromofluoromethane	8.022	113	173283	50.585	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.160%
50) Toluene-d8	10.439	98	581934	43.046	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	86.100%
62) 4-Bromofluorobenzene	12.727	95	228386	50.327	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.660%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	41342	14.599	ug/l	96
20) Methylene Chloride	5.092	84	29692	6.674	ug/l	93
43) Isopropyl Acetate	8.557	43	51392	4.519	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072849.D
Acq On : 14 Jun 2022 23:03
Operator : JC\MD
Sample : PB145494ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145494ZHE#01

Quant Time: Jun 15 04:36:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

