

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072850.D
 Acq On : 14 Jun 2022 23:27
 Operator : JC\MD
 Sample : PB145494ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145494ZHE#02

Quant Time: Jun 15 04:36:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	317343	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	538103	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	476777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	175984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	206667	42.049	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.100%
35) Dibromofluoromethane	8.016	113	169905	50.999	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.000%
50) Toluene-d8	10.439	98	574287	43.680	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.360%
62) 4-Bromofluorobenzene	12.727	95	213076	48.279	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.560%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	40048	14.579	ug/l	90
20) Methylene Chloride	5.093	84	30338	7.131	ug/l	91
43) Isopropyl Acetate	8.563	43	56330	5.093	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072850.D
Acq On : 14 Jun 2022 23:27
Operator : JC\MD
Sample : PB145494ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145494ZHE#02

Quant Time: Jun 15 04:36:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

