

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072852.D
 Acq On : 15 Jun 2022 00:16
 Operator : JC\MD
 Sample : PB145494ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145494ZHE#04

Quant Time: Jun 15 06:36:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	321717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	558234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	513337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	210228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	213623	42.873	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.740%
35) Dibromofluoromethane	8.016	113	173228	50.121	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.240%
50) Toluene-d8	10.439	98	601320	44.086	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.180%
62) 4-Bromofluorobenzene	12.727	95	244892	53.487	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.980%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	39574	14.211	ug/l	93
20) Methylene Chloride	5.098	84	28272	6.403	ug/l	87
43) Isopropyl Acetate	8.563	43	68938	6.008	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072852.D
Acq On : 15 Jun 2022 00:16
Operator : JC\MD
Sample : PB145494ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145494ZHE#04

Quant Time: Jun 15 06:36:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

