

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072853.D
 Acq On : 15 Jun 2022 00:40
 Operator : JC\MD
 Sample : PB145494ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145494ZHE#05

Quant Time: Jun 15 04:37:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	315563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	536080	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	475426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	184104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	207639	42.485	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.980%
35) Dibromofluoromethane	8.016	113	168540	50.780	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.560%
50) Toluene-d8	10.439	98	568574	43.408	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	86.820%
62) 4-Bromofluorobenzene	12.727	95	220565	50.164	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.320%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	39164	14.338	ug/l #	86
20) Methylene Chloride	5.093	84	29575	6.954	ug/l #	84
43) Isopropyl Acetate	8.563	43	68418	6.209	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072853.D
Acq On : 15 Jun 2022 00:40
Operator : JC\MD
Sample : PB145494ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145494ZHE#05

Quant Time: Jun 15 04:37:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

