

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072856.D
 Acq On : 15 Jun 2022 01:53
 Operator : JC\MD
 Sample : PB145494ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145494ZHE#08

Quant Time: Jun 15 04:37:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	312561	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	539706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	487565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	192155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	208201	43.009	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.020%
35) Dibromofluoromethane	8.022	113	168864	50.536	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.080%
50) Toluene-d8	10.439	98	579776	43.966	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.940%
62) 4-Bromofluorobenzene	12.727	95	229667	51.883	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.760%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	30194	11.160	ug/l	94
20) Methylene Chloride	5.093	84	16140	2.986	ug/l	93
43) Isopropyl Acetate	8.557	43	91523	8.250	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072856.D
Acq On : 15 Jun 2022 01:53
Operator : JC\MD
Sample : PB145494ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145494ZHE#08

Quant Time: Jun 15 04:37:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

