

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072861.D
 Acq On : 15 Jun 2022 03:55
 Operator : JC\MD
 Sample : N3329-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PALMYRA-RT-73-COMP-2

Quant Time: Jun 15 04:38:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	295397	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	513515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	459756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	169956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	198852	43.465	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.920%
35) Dibromofluoromethane	8.022	113	160182	50.383	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.760%
50) Toluene-d8	10.439	98	550113	43.844	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.680%
62) 4-Bromofluorobenzene	12.727	95	209518	49.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.500%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	52956	20.711	ug/l	90
20) Methylene Chloride	5.093	84	85496	25.405	ug/l	86
25) 2-Butanone	7.339	43	23069	5.694	ug/l	92
43) Isopropyl Acetate	8.569	43	83952	7.954	ug/l	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072861.D
Acq On : 15 Jun 2022 03:55
Operator : JC\MD
Sample : N3329-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PALMYRA-RT-73-COMP-2

Quant Time: Jun 15 04:38:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

