

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072862.D
 Acq On : 15 Jun 2022 04:20
 Operator : JC\MD
 Sample : N3329-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PALMYRA-RT-73-COMP-3

Quant Time: Jun 15 07:51:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 15 04:13:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	301108	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	516818	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	461378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	178430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	202626	43.450	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.900%
35) Dibromofluoromethane	8.016	113	161217	50.384	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.760%
50) Toluene-d8	10.439	98	552360	43.742	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.480%
62) 4-Bromofluorobenzene	12.727	95	216908	51.171	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.340%
Target Compounds						
					Qvalue	
16) Acetone	4.286	43	51580	19.790	ug/l	99
20) Methylene Chloride	5.098	84	88673	25.882	ug/l #	83
25) 2-Butanone	7.333	43	23979	5.806	ug/l #	78
43) Isopropyl Acetate	8.557	43	83825	7.891	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
Data File : VN072862.D
Acq On : 15 Jun 2022 04:20
Operator : JC\MD
Sample : N3329-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PALMYRA-RT-73-COMP-3

Quant Time: Jun 15 07:51:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 15 04:13:32 2022
Response via : Initial Calibration

