



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061422\  
Data File : VN072827.D  
Acq On : 14 Jun 2022 10:37  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061422W.M  
Title : SW846 8260  
Last Update : Wed Jun 15 04:13:32 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 20.7         | 10462      | PASS                |
| 75             | 95              | 30              | 60              | 52.9         | 26797      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 50656      | PASS                |
| 96             | 95              | 5               | 9               | 6.6          | 3338       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.4          | 506        | PASS                |
| 174            | 95              | 50              | 100             | 70.2         | 35555      | PASS                |
| 175            | 174             | 5               | 9               | 7.3          | 2605       | PASS                |
| 176            | 174             | 95              | 101             | 98.2         | 34923      | PASS                |
| 177            | 176             | 5               | 9               | 6.7          | 2349       | PASS                |