



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061422\
 Data File : VN072836.D
 Acq On : 14 Jun 2022 16:12
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061422W.M
 Title : SW846 8260
 Last Update : Wed Jun 15 04:13:32 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	11061	PASS
75	95	30	60	51.1	27269	PASS
95	95	100	100	100.0	53325	PASS
96	95	5	9	6.8	3630	PASS
173	174	0.00	2	0.9	327	PASS
174	95	50	100	68.0	36259	PASS
175	174	5	9	8.0	2914	PASS
176	174	95	101	98.6	35739	PASS
177	176	5	9	7.2	2576	PASS