

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078194.D
 Acq On : 14 Jun 2023 10:19
 Operator : JC\MD
 Sample : VN0614WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0614WBL01

Quant Time: Jun 15 02:39:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	523733	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	955630	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	833278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	257275	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	364870	52.251	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.500%
35) Dibromofluoromethane	8.171	113	325781	53.917	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.840%
50) Toluene-d8	10.571	98	1109242	47.620	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.240%
62) 4-Bromofluorobenzene	12.853	95	350774	48.440	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078194.D
Acq On : 14 Jun 2023 10:19
Operator : JC\MD
Sample : VN0614WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0614WBL01

Quant Time: Jun 15 02:39:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

