

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078197.D
 Acq On : 14 Jun 2023 11:44
 Operator : JC\MD
 Sample : 03225-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-061323

Quant Time: Jun 15 02:41:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	471464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	867055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	741050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	218757	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	324778	51.666	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.340%
35) Dibromofluoromethane	8.177	113	293918	53.612	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.220%
50) Toluene-d8	10.570	98	1008839	47.734	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.460%
62) 4-Bromofluorobenzene	12.853	95	308388	46.937	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.880%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	9467	5.232	ug/l	88
20) Methylene Chloride	5.294	84	8269	1.383	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078197.D
Acq On : 14 Jun 2023 11:44
Operator : JC\MD
Sample : 03225-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-061323

Quant Time: Jun 15 02:41:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

