

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078201.D
 Acq On : 14 Jun 2023 13:20
 Operator : JC\MD
 Sample : 03225-03DL 4X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-05-84-104-061323DL

Quant Time: Jun 15 02:42:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	471849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	876645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	766292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	234412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	332320	52.823	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.640%	
35) Dibromofluoromethane	8.171	113	296785	53.543	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.080%	
50) Toluene-d8	10.576	98	1029663	48.186	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.380%	
62) 4-Bromofluorobenzene	12.853	95	321151	48.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.680%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	4.353	96	115962	22.506	ug/l #	81
21) trans-1,2-Dichloroethene	5.794	96	10208	1.837	ug/l	87
24) 1,1-Dichloroethane	6.588	63	14801	1.517	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	84379	12.729	ug/l	97
42) 1,2-Dichloroethane	8.682	62	9797	1.204	ug/l	94
44) Trichloroethene	9.359	130	84527	13.485	ug/l	97
49) 1,4-Dioxane	9.712	88	2007	19.156	ug/l #	90
52) Toluene	10.641	92	3522	0.235	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078201.D
Acq On : 14 Jun 2023 13:20
Operator : JC\MD
Sample : 03225-03DL 4X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-05-84-104-061323DL

Quant Time: Jun 15 02:42:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

