

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078203.D
 Acq On : 14 Jun 2023 14:08
 Operator : JC\MD
 Sample : 03160-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PNNL-B-0609-1

Quant Time: Jun 15 02:43:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	452692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	844693	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	773080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	242729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	324267	53.724	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.440%		
35) Dibromofluoromethane	8.171	113	293896	55.028	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.060%		
50) Toluene-d8	10.571	98	1021990	49.636	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.280%		
62) 4-Bromofluorobenzene	12.853	95	323225	50.497	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.000%		
Target Compounds						
16) Acetone	4.453	43	25686	14.785	ug/l #	84
20) Methylene Chloride	5.283	84	20523	3.575	ug/l	98
37) Ethyl Acetate	7.577	43	55387	9.255	ug/l #	76
43) Isopropyl Acetate	8.694	43	230041	20.099	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078203.D
Acq On : 14 Jun 2023 14:08
Operator : JC\MD
Sample : 03160-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PNNL-B-0609-1

Quant Time: Jun 15 02:43:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

