

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078209.D
 Acq On : 14 Jun 2023 16:32
 Operator : JC\MD
 Sample : PB153393TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153393TB

Quant Time: Jun 15 02:46:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	464400	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	846272	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	799289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	249193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	327852	52.949	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.900%
35) Dibromofluoromethane	8.171	113	294513	55.040	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.080%
50) Toluene-d8	10.571	98	1031281	49.994	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.980%
62) 4-Bromofluorobenzene	12.853	95	325121	50.699	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.400%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	26088	14.638	ug/l #	86
20) Methylene Chloride	5.289	84	16389	2.783	ug/l #	89
37) Ethyl Acetate	7.577	43	58702	9.790	ug/l	98
43) Isopropyl Acetate	8.700	43	285770	24.921	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078209.D
Acq On : 14 Jun 2023 16:32
Operator : JC\MD
Sample : PB153393TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153393TB

Quant Time: Jun 15 02:46:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

