

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078210.D
 Acq On : 14 Jun 2023 16:57
 Operator : JC\MD
 Sample : 03157-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-622-VOC-01

Quant Time: Jun 15 02:46:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	467034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	876554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	798423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	258489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	326793	52.480	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.960%		
35) Dibromofluoromethane	8.177	113	292100	52.703	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.400%		
50) Toluene-d8	10.571	98	1049836	49.135	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.280%		
62) 4-Bromofluorobenzene	12.853	95	333793	50.253	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.500%		
Target Compounds						
16) Acetone	4.442	43	481481	268.630	ug/l	Qvalue 99
20) Methylene Chloride	5.289	84	27180	4.589	ug/l #	84
37) Ethyl Acetate	7.571	43	47030	7.573	ug/l	97
43) Isopropyl Acetate	8.694	43	254444	21.423	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078210.D
Acq On : 14 Jun 2023 16:57
Operator : JC\MD
Sample : 03157-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-622-VOC-01

Quant Time: Jun 15 02:46:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

