

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078214.D
 Acq On : 14 Jun 2023 18:33
 Operator : JC\MD
 Sample : 03231-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-2D-06132023

Quant Time: Jun 15 02:47:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	444944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	825006	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	732854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	235051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	321450	54.185	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.360%			
35) Dibromofluoromethane	8.177	113	285165	54.667	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.340%			
50) Toluene-d8	10.576	98	972635	48.366	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.740%			
62) 4-Bromofluorobenzene	12.853	95	300217	48.022	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.040%			
Target Compounds						Qvalue
38) Carbon Tetrachloride	8.382	117	4061	0.514	ug/l #	70
44) Trichloroethene	9.365	130	6671	1.131	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078214.D
Acq On : 14 Jun 2023 18:33
Operator : JC\MD
Sample : 03231-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-2D-06132023

Quant Time: Jun 15 02:47:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

