

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078216.D
 Acq On : 14 Jun 2023 19:21
 Operator : JC\MD
 Sample : 03231-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-3S-06122023

Quant Time: Jun 15 02:48:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	439387	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	821714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	719056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	226898	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	313965	53.593	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.180%	
35) Dibromofluoromethane	8.171	113	276359	53.191	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.380%	
50) Toluene-d8	10.576	98	966872	48.272	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.540%	
62) 4-Bromofluorobenzene	12.853	95	297018	47.701	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.400%	
Target Compounds						
					Qvalue	
44) Trichloroethene	9.359	130	396353	67.457	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078216.D
Acq On : 14 Jun 2023 19:21
Operator : JC\MD
Sample : 03231-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-3S-06122023

Quant Time: Jun 15 02:48:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

