

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
 Data File : VN078217.D
 Acq On : 14 Jun 2023 19:45
 Operator : JC\MD
 Sample : 03231-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-1D-06132023

Quant Time: Jun 15 02:48:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/15/2023
 Supervised By : Mahesh Dadoda 06/15/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	435579	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	822189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	720060	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	226189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	318012	54.758	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.520%
35) Dibromofluoromethane	8.177	113	284910	54.805	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.620%
50) Toluene-d8	10.570	98	976935	48.747	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.500%
62) 4-Bromofluorobenzene	12.853	95	298086	47.845	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.680%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	4.341	96	5023m	1.056	ug/l	
21) trans-1,2-Dichloroethene	5.788	96	2120m	0.413	ug/l	
27) cis-1,2-Dichloroethene	7.500	96	14777	2.415	ug/l	96
44) Trichloroethene	9.359	130	1487049	252.941	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061423\
Data File : VN078217.D
Acq On : 14 Jun 2023 19:45
Operator : JC\MD
Sample : 03231-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-1D-06132023

Quant Time: Jun 15 02:48:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/15/2023
Supervised By : Mahesh Dadoda 06/15/2023

